CONNECTING RODS INSTRUCTION GUIDE



STEP-BY-STEP DIAGRAM:


Based on the starting protrusive bite position, the initial connecting rods will be pre-selected and placed on the appliance.






3 Rods have Left and Right listed on them as








CHANGING RODS


The connecting rods can be titrated in 1mm increments based on the numbering system listed on the card and on the rods.



The rods are removed from the holding bar by twisting. If a small bur remains on the rod it can be easily trimmed with a scissor.

INSERTING ROD

When inserting new connecting rods, make sure that the numbers are facing lingually toward the tongue, not toward the cheek.



SML® - Space Maintainers Laboratories 9129 Lurline Ave, Chatsworth, CA 1-800-423-3270 www.SMLglobal.com